


HIOKI

LCR testiranje u istraživanju i razvoju

	<p>ANALIZATOR IMPEDANSE IM7580</p> <ul style="list-style-type: none">• Z, L, C, R testiranje• 1 MHz do 300 MHz izvorna test frekvencija• Maksimum test brzina od 0.5 ms• Meri LCR i simultano izvodi frekvencijska prebrisavanja• Koristi se i pri testiranju tokom proizvodnje komponenti <p>Više na: http://www.hioki.com/products/lcr_resistance_signal/lcr_resistance_meters/1500</p>
	<p>ANALIZATOR IMPEDANSE IM7581</p> <ul style="list-style-type: none">• Z, L, C, R testiranje• 100 kHz do 300 MHz izvorna test frekvencija• Maksimum test brzina od 0.5 ms• Meri LCR i simultano izvodi frekvencijska prebrisavanja• Promenljivost merene vrednosti: 0,07% <p>Više na: https://www.hioki.com/en/products/detail/?product_key=6319</p>
	<p>ANALIZATOR HEMIJSKE IMPEDANSE IM3590</p> <ul style="list-style-type: none">• Z, L, C, R, σ, ϵ testiranje• Merenja baterija• 1 mHz do 200 kHz• 2 ms merno vreme <p>Više na: http://www.hioki.com/products/lcr_resistance_signal/lcr_resistance_meters/713</p>
	<p>LCR METAR IM3533, IM3533-01</p> <ul style="list-style-type: none">• Z, L, C, R testiranje• 1 mHz do 200 kHz• 2 ms merno vreme• Transformator režim• Frekvencijsko prebrisavanje w/IM3533-01 <p>Više na: http://www.hioki.com/products/lcr_resistance_signal/lcr_resistance_meters/653</p>
	<p>ANALIZATOR IMPEDANSE IM3570</p> <ul style="list-style-type: none">• Z, L, C, R testiranje• 4 Hz do 5 MHz• 0.5 ms merno vreme• Meri LCR i simultano izvodi frekvencijska prebrisavanja <p>Više na: http://www.hioki.com/products/lcr_resistance_signal/lcr_resistance_meters/654</p>
	<p>LCR TESTER 3535</p> <ul style="list-style-type: none">• Z, L, C, R testiranje• 100kHz do 120MHz• 6 ms merno vreme <p>Više na: http://www.hioki.com/products/lcr_resistance_signal/lcr_resistance_meters/541</p>

	<p>LCR TESTER 3532-50</p> <ul style="list-style-type: none"> • Z , L, C, R testiranje • 42Hz do 5MHz • 5 ms merno vreme <p>Više na: http://www.hioki.com/products/lcr_resistance_signal/lcr_resistance_meters/539</p>
	<p>ANALIZATORI IMPEDANSE IM7583 i IM7535</p> <ul style="list-style-type: none"> • Osnovna tačnost: Z: +/-0.65% rdg. theta: +/-0.38deg. (reprezentativne vrednosti) • Merno vreme: Kratko poput 0.5 ms (analogno merenje) • Merni opseg: 100 mOhm do 5kOhm • Merne frekvencije: IM7583: 1 MHz do 600 MHz / IM7585: 1 MHz do 1.3 GHz • Nivo mernog signala: -40.0 dBm do +1.0 dBm (4 mV do 502 mV) <p>Više na: http://www.hioki.com/products/lcr_resistance_signal/lcr_resistance_meters/</p>

Testiranje LCR i kapacitivnosti tokom proizvodnje komponenti

	<p>C METAR 3506-10</p> <ul style="list-style-type: none"> • C, D (tan δ), Q, i nisko kapacitivno testiranje • 1 kHz, 1 MHz • 1.5 ms merno vreme pri 1 MHz <p>Više na: http://www.hioki.com/products/lcr_resistance_signal/lcr_resistance_meters/1399</p>
	<p>LCR METAR IM3523</p> <ul style="list-style-type: none"> • Z , L, C, R testiranje • 40 Hz do 200 kHz • 2 ms merno vreme <p>Više na: http://www.hioki.com/products/lcr_resistance_signal/lcr_resistance_meters/652</p>
	<p>C TESTER 3504 Serija</p> <ul style="list-style-type: none"> • C, D (tan δ), visokokapacitivno kapacitivna MLCC testiranje • 120Hz ili 1kHz • 2 ms merno vreme <p>Više na: http://www.hioki.com/products/lcr_resistance_signal/lcr_resistance_meters/532</p>
	<p>LCR TESTER 3511-50</p> <ul style="list-style-type: none"> • Z , L, C, R testiranje • 120Hz ili 1kHz • 5 ms merno vreme <p>Više na: http://www.hioki.com/products/lcr_resistance_signal/lcr_resistance_meters/537</p>

Melco Buda d.o.o.

- kancelarija u Beogradu: Hadži Nikole Živkovića br.2
 Poslovna zgrada Iskra komerc, kancelarija 15/ II sprat
 tel: 011/ 2181 609, SBB tel: 011/40-55-420, tel/faks: 011/ 3286 445
 e mail: office-beograd@melcobuda.co.rs , budimir.melcobuda@gmail.com
www.melcobuda.co.rs , www.kyoritsu-instrumenti.com , www.temovizija.com , www.hioki-instrumenti.com

- kancelarija u Despotovcu: Saveza Boraca br.7, 35213 Despotovac, Srbija
 tel:035/612 916, faks:035/613 319, mob. 063/8003370
 e mail: office@kyoritsu-instrumenti.com , office@melcobuda.co.rs

- Germany address: Quer strasse 18 Offenbach